

NABL ACCREDITATION | CW-408 | HEPA FILTER

National Accreditation Board for Testing and Calibration Laboratories, Department of Science and Technology, Government of India.



Parashar Micro Measurement Pvt. Ltd.
B-59, SECTOR-64, NODA (U.P.) 201 307
PH: 0120-4252360-65, Email: pmmppl.india@gmail.com
Website: www.pmmppl.in, www.parashar.info

Calibration Certificate

Page 1 of 2

Certificate Number: PMM/231124/11-2	Date of Request: 23.11.2024
Calibrated For: M/s. Anix Academy Of Data Sciences. (Froyo Technologies) 102, 1st Floor Bhagimtal Complex Naya Bank, Sector - 15, Noida, U.P. - 201301	Calibrated on: 23.11.2024 Suggested Due Date: 22.11.2025 Cert. Issue Date: 27.11.2024

Description of Equipments	
Nomenclature: Hepa Filter (Portable) Range: Refer to Result Accuracy/Class: Not Mentioned Condition of UUC: Physically OK Master Equipment / Standard Used: Party ID Mark No. CN - 408 Location:	Serial No. / ID Mark: _____ Calibrated From: _____ Certificate No.: _____ Due Date of Calibration: _____

Sr. No.	Nomenclature	Make & Model	Serial No. / ID Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Airborne Particle Counter	Bidbase / CL-1003	CLJ202324020	SPM Lab Solution	SPM/N/2024/093	26.05.2025
2	Aerosoliser	Sigma	PMM/S/AM/02	Aadon Gurgaon	ATCL24-254/020	17.02.2025

Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC: 17025-2017 or National Standards through unbroken chain of calibration.

Calibration Performed at	Site	Environmental Condition(s)	Temperature = 23 ± 0.5 °C	Humidity = 30 to 75 % RH
Reference Standard / Guideline: PMM/TCP/01	ISO 14644-1			

ACCEPTANCE CRITERIA	
ISO Cleanroom Classes	Particle Counter per m ³ as per ISO: 14644-1
≥ 0.3 µm	≥ 1.0 µm
≥ 0.5 µm	≥ 3.0 µm
10	35
ISO CLASS 3	352
ISO CLASS 4	83
ISO CLASS 5	832
ISO CLASS 6	35200
ISO CLASS 7	83200
ISO CLASS 8	352000

Calibrated By: Suraj Maurya
Designation: Calibration Engineer

Approved By: Jai Prakash
Designation: Tech Manager / Auth. Sign.

Conditions: 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of Chief Executive Officer of the lab.

Parashar Micro Measurement Pvt. Ltd.
B-59, SECTOR-64, NODA (U.P.) 201 307
PH: 0120-4252360-65, Email: pmmppl.india@gmail.com
Website: www.pmmppl.in, www.parashar.info

Calibration Certificate

Page 2 of 2

Dust Particle Count & Hepa Filter Integrity (Status Report)			
S. No.	Outlet Particle Count at m ³ (Corner A)	Outlet Particle Count at m ³ (Corner B)	Outlet Particle Count at m ³ (Centre C)
1	17:35:42	17:45:43	17:55:44
2	18:05:45	18:20:46	

TIME	PERIOD	VOLUME	SIZE	CLIMU	CLIMU	CLIMU
18:05:45	60 Sec	1 m ³	0.3 µm	5243	5377	5377
18:20:46	60 Sec	1 m ³	0.3 µm	1852	1482	212
			0.5 µm	220	0	0
			1.0 µm	0	0	0
			3.0 µm	0	0	0
			5.0 µm	0	0	0
			10.0 µm	0	0	0

Note: Results within ISO CLASS 5 as per ISO 14644-1 STANDARD

RESULTS OF AIR FLOW	
S.N.	Air Flow in FPM
1	104
2	109
3	114
4	111
5	109
Mean	
109.40	

UUC - Unit Under Calibration
Std. - Standard Instrument
Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor k = 2) = ± 2% of FSD

Calibrated By: Suraj Maurya
Designation: Calibration Engineer

Approved By: Jai Prakash
Designation: Tech Manager / Auth. Sign.

Conditions: 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of Chief Executive Officer of the lab.

ISO 14644-1:2015 specifies the classification of air cleanliness in terms of concentration of airborne particles in cleanrooms and clean zones; and separative devices as defined in ISO 14644-7.